

**Great Plains Industries, Inc.**  
 5252 E. 36th St. N.  
 Wichita, KS 67220  
**Tel:** 888-996-3837  
**Fax:** 316-686-6746  
**Email:** [meters@gplains.com](mailto:meters@gplains.com)



"Look for the blue label!"

## G2 Stainless Steel Industrial Flowmeters

The GPI Stainless Steel Flowmeter line are used with ammonium, plating solutions, caustic and acidic cleaning agents, fuel products and other chemicals. The lithium battery powered electronic display G2 Stainless Steel flowmeters are calibrated to read in gallons and liters per minute with two totals in which case one is resettable and gives you a cumulative flow total. The G2 Stainless Steel flowmeters come in ½ in, ¾ in, 1 in, 1½ in, and 2 inch models.

Watch the G2 Stainless Steel video on YouTube.

Watch How To Calibrate an 09 Computer Display on YouTube.

Watch How To Convert your 09 Computer Display from Local to Remote on YouTube.  
 FindADistributor();

Results 1 - 140 of 140

Item #	Inner Diameter	Fitting	Accuracy	Flow Range	Display	Reads	Typical K-Factor	Popularity
G2S05N09GMA	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	<b>Most Popular</b>
G2S05N19GMA	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	N/A
G2S05N43GMC	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	N/A
G2S05N53GMC	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	N/A
G2S05N63GMC	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	N/A
G2S05N73GMC	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	N/A
G2S05N41XXC	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	No Display	No Calibration	2,500	N/A

G2S05N51GMC	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - GG500 Transmitter	Gallons Litres	2,500	N/A
G2S05N52GMC	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - GG500 Transmitter	Gallons Litres	2,500	N/A
G2S05N61GMC	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - GX500 Transmitter	Gallons Litres	2,500	N/A
G2S05N62GMC	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - GX500 Transmitter	Gallons Litres	2,500	N/A
G2S05N71XXC	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	No Display	No Calibration	2,500	N/A
G2S05N72XXC	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	No Display	No Calibration	2,500	N/A
G2S05NXXXXA	½ in.	NPT ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	No Display	Turbine Only	2,500	N/A
G2S05I09LMA	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	N/A
G2S05I19LMA	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	N/A
G2S05I43LMC	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	N/A
G2S05I53LMC	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	N/A
G2S05I63LMC	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	N/A

G2S05173LMC	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - 09 Computer	Gallons Litres	2,500	N/A
G2S05141XXC	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	No Display	No Calibration	2,500	N/A
G2S05151LMC	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - GG500 Transmitter	Gallons Litres	2,500	N/A
G2S05152LMC	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - GG500 Transmitter	Gallons Litres	2,500	N/A
G2S05161LMC	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - GX500 Transmitter	Gallons Litres	2,500	N/A
G2S05162LMC	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	LCD - GX500 Transmitter	Gallons Litres	2,500	N/A
G2S05171XXC	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	No Display	No Calibration	2,500	N/A
G2S05172XXC	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	No Display	No Calibration	2,500	N/A
G2S051XXXXA	½ in.	ISO ½ in.	±1.5 % of reading with Computer ±2.0 % of reading without Computer	1 to 10 GPM 3.8 to 37.9 LPM	No Display	Turbine Only	2,500	N/A
G2S07N09GMA	¾ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	<b>Most Popular</b>
G2S07N19GMA	¾ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	N/A
G2S07N43GMC	¾ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	N/A

G2S07N53GMC	¼ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	N/A
G2S07N63GMC	¼ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	N/A
G2S07N73GMC	¼ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	N/A
G2S07N41XXC	¼ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	No Display	No Calibration	1,100	N/A
G2S07N51GMC	¼ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - GG500 Transmitter	Gallons Litres	1,100	N/A
G2S07N52GMC	¼ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - GG500 Transmitter	Gallons Litres	1,100	N/A
G2S07N61GMC	¼ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - GX500 Transmitter	Gallons Litres	1,100	N/A
G2S07N62GMC	¼ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - GX500 Transmitter	Gallons Litres	1,100	N/A
G2S07N71XXC	¼ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	No Display	No Calibration	1,100	N/A
G2S07N72XXC	¼ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	No Display	No Calibration	1,100	N/A
G2S07NXXXXA	¼ in.	NPT ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	No Display	Turbine Only	1,100	N/A
G2S07I09LMA	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	N/A

G2S07119LMA	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	N/A
G2S07143LMC	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	N/A
G2S07153LMC	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	N/A
G2S07163LMC	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	N/A
G2S07173LMC	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - 09 Computer	Gallons Litres	1,100	N/A
G2S07141XXC	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	No Display	No Calibration	1,100	N/A
G2S07151LMC	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - GG500 Transmitter	Gallons Litres	1,100	N/A
G2S07152LMC	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - GG500 Transmitter	Gallons Litres	1,100	N/A
G2S07161LMC	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - GX500 Transmitter	Gallons Litres	1,100	N/A
G2S07162LMC	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	LCD - GX500 Transmitter	Gallons Litres	1,100	N/A
G2S07171XXC	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	No Display	No Calibration	1,100	N/A
G2S07172XXC	¼ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	No Display	No Calibration	1,100	N/A

G2S071XXXXA	¾ in.	ISO ¾ in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	2 to 20 GPM 7.6 to 75.7 LPM	No Display	Turbine Only	1,100	N/A
G2S10N09GMA	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	<b>Most Popular</b>
G2S10N19GMA	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	N/A
G2S10N43GMC	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	N/A
G2S10N53GMC	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	N/A
G2S10N63GMC	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	N/A
G2S10N73GMC	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	N/A
G2S10N41XXC	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	No Display	No Calibration	565	N/A
G2S10N51GMC	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - GG500 Transmitter	Gallons Litres	565	N/A
G2S10N52GMC	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - GG500 Transmitter	Gallons Litres	565	N/A
G2S10N61GMC	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - GX500 Transmitter	Gallons Litres	565	N/A
G2S10N62GMC	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - GX500 Transmitter	Gallons Litres	565	N/A

G2S10N71XXC	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	No Display	No Calibration	565	N/A
G2S10N72XXC	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	No Display	No Calibration	565	N/A
G2S10NXXXXA	1 in.	NPT 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	No Display	Turbine Only	565	N/A
G2S10I09LMA	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	N/A
G2S10I19LMA	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	N/A
G2S10I43LMC	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	N/A
G2S10I53LMC	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	N/A
G2S10I63LMC	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	N/A
G2S10I73LMC	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - 09 Computer	Gallons Litres	565	N/A
G2S10I41XXC	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	No Display	No Calibration	565	N/A
G2S10I51LMC	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - GG500 Transmitter	Gallons Litres	565	N/A
G2S10I52LMC	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - GG500 Transmitter	Gallons Litres	565	N/A

G2S10I61LMC	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - GX500 Transmitter	Gallons Litres	565	N/A
G2S10I62LMC	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	LCD - GX500 Transmitter	Gallons Litres	565	N/A
G2S10I71XXC	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	No Display	No Calibration	565	N/A
G2S10I72XXC	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	No Display	No Calibration	565	N/A
G2S10IXXXXA	1 in.	ISO 1 in.	±1.0 % of reading with Computer ±1.5 % of reading without Computer	5 to 50 GPM 18.9 to 190 LPM	No Display	Turbine Only	565	N/A
G2S15N09GMB	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	<b>Most Popular</b>
G2S15N19GMB	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	N/A
G2S15N43GMC	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	N/A
G2S15N53GMC	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	N/A
G2S15N63GMC	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	N/A
G2S15N73GMC	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	N/A
G2S15N41XXC	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	No Display	No Calibration	215	N/A



G2S15N51GMC	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - GG500 Transmitter	Gallons Litres	215	N/A
G2S15N52GMC	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - GG500 Transmitter	Gallons Litres	215	N/A
G2S15N61GMC	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - GX500 Transmitter	Gallons Litres	215	N/A
G2S15N62GMC	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - GX500 Transmitter	Gallons Litres	215	N/A
G2S15N71XXC	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	No Display	No Calibration	215	N/A
G2S15N72XXC	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	No Display	No Calibration	215	N/A
G2S15NXXXXB	1 ½ in.	NPT 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	No Display	Turbine Only	215	N/A
G2S15I09LMB	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	N/A
G2S15I19LMB	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	N/A
G2S15I43LMC	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	N/A
G2S15I53LMC	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	N/A
G2S15I63LMC	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	N/A

G2S15173LMC	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - 09 Computer	Gallons Litres	215	N/A
G2S15141XXC	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	No Display	No Calibration	215	N/A
G2S15151LMC	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - GG500 Transmitter	Gallons Litres	215	N/A
G2S15152LMC	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - GG500 Transmitter	Gallons Litres	215	N/A
G2S15161LMC	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - GX500 Transmitter	Gallons Litres	215	N/A
G2S15162LMC	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	LCD - GX500 Transmitter	Gallons Litres	215	N/A
G2S15171XXC	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	No Display	No Calibration	215	N/A
G2S15172XXC	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	No Display	No Calibration	215	N/A
G2S151XXXXB	1 ½ in.	ISO 1 ½ in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	10 to 100 GPM 38.0 to 380 LPM	No Display	Turbine Only	215	N/A
G2S20N09GMB	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	<b>Most Popular</b>
G2S20N19GMB	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	N/A
G2S20N43GMC	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	N/A

G2S20N53GMC	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	N/A
G2S20N63GMC	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	N/A
G2S20N73GMC	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	N/A
G2S20N41XXC	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	No Display	No Calibration	100	N/A
G2S20N51GMC	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - GG500 Transmitter	Gallons Litres	100	N/A
G2S20N52GMC	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - GG500 Transmitter	Gallons Litres	100	N/A
G2S20N61GMC	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - GX500 Transmitter	Gallons Litres	100	N/A
G2S20N62GMC	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - GX500 Transmitter	Gallons Litres	100	N/A
G2S20N71XXC	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	No Display	No Calibration	100	N/A
G2S20N72XXC	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	No Display	No Calibration	100	N/A
G2S20NXXXXB	2 in.	NPT 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	No Display	Turbine Only	100	N/A
G2S20I09LMB	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	N/A

G2S20I19LMB	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	N/A
G2S20I43LMC	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	N/A
G2S20I53LMC	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	N/A
G2S20I63LMC	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	N/A
G2S20I73LMC	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - 09 Computer	Gallons Litres	100	N/A
G2S20I41XXC	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	No Display	No Calibration	100	N/A
G2S20I51LMC	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - GG500 Transmitter	Gallons Litres	100	N/A
G2S20I52LMC	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - GG500 Transmitter	Gallons Litres	100	N/A
G2S20I61LMC	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - GX500 Transmitter	Gallons Litres	100	N/A
G2S20I62LMC	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	LCD - GX500 Transmitter	Gallons Litres	100	N/A
G2S20I71XXC	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	No Display	No Calibration	100	N/A
G2S20I72XXC	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	No Display	No Calibration	100	N/A

G2S20XXXXB	2 in.	ISO 2 in.	±0.75 % of reading with Computer ±1.0 % of reading without Computer	20 to 200 GPM 76 to 760 LPM	No Display	Turbine Only	100	N/A
------------	-------	--------------	--	--------------------------------	------------	--------------	-----	-----

---

Results 1 - 140 of 140